Medix Biochemica

Anti-h sFlt-1 NA000402 SPTN-5

| Product overview | | | | | |
|----------------------------|--|--|--|--|--|
| Catalog number | C-10-0012 | | | | |
| Specificity | Antibody recognizes human sFlt-1 (soluble fms-like tyrosine kinase-1) | | | | |
| Description | Monoclonal mouse antibody, cultured in vitro under conditions free from animal-derived components. | | | | |
| Product buffer solution | 50 mM Na-citrate, pH 6.0, 0.9 % NaCI, 0.1% ProClin [™] 300 as a preservative | | | | |
| Shelf life and storage | Unspecified, storage at 2–8 °C | | | | |
| Subclass | IgG1 | | | | |
| Analyte description | sFlt-1 was originally found in human umbilical vein endothelial cells. It is a glycoprotein with tyrosine kinase activity, which is formed by different splicing of VEGFR-1 mRNA. It has 7 isoforms. Abnormally high expression of soluble isoforms (isoform 2, isoform 3 or isoform 4) may be a cause of preeclampsia. sFlt-1 is always used in combination with PIGF to determine the sFlt-1/PIGF ratio. The sFlt-1/PIGF ratio is intended for use as an aid in the diagnosis of preeclampsia in conjunction with other diagnostic and clinical information. | | | | |
| Parameters tested on ea | ch lot | | | | |
| Product appearance | Liquid, may turn slightly opaque during storage | | | | |
| Product concentration | 5.0 mg/ml (+/- 10%) | | | | |
| Immunoreactivity | 80–120% compared to the reference sample in an ELISA test | | | | |
| IEF Profile | | | | | |
| Purity | ≥ 95 % | | | | |
| Kinetic parameters | | | | | |
| Association rate constant | To Be Determined (TBD) | | | | |
| Dissociation rate constant | TBD | | | | |
| Affinity constant | TBD | | | | |
| Determination method | - | | | | |
| Determination antigen | - | | | | |



Medix Biochemica China 上海劢迪克斯生物化学有限公司 上海市闵行区浦江绿洲环路10号6幢11层

网址: www.cnmedixbiochemica.com

Medix Biochemica

Cross-reactivities

Epitope

Not Determined (N/D)

| Pair rec | ommen | dations |
|----------|-------|---------|
|----------|-------|---------|

| | | DETECTION | | | |
|---------|----------|-----------|----------|----------|----------|
| | | NA000401 | NA000402 | NA000403 | NA000404 |
| CAPTURE | NA000401 | - | - | + | - |
| | NA000402 | - | - | + | - |
| | NA000403 | + | + | - | + |
| | NA000404 | + | - | + | - |

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

| Platforms tested | ELISA, CLIA | |
|-------------------|---|--|
| Antigens tested | - | |
| Product stability | TEMPERATURE, TIME -70 °C, 21 days -20 °C, 21 days +4 °C, 21 days +35 °C, 21 days +45 °C, 7 days Stability testing is performed in the p temperatures affect the antigen bindi antibody. Please note that the shelf li real time stability testing at 2–8 °C in | fe given on the first page is based on |
| Miscellaneous | - | |
| References | - | |

